


# Technical data

## OCULUS C-Quant

Header	
Light source	White LED
Maximum brightness	300 cd/m <sup>2</sup>
Size of test field	14°
Examination duration	ca. 1 minute
Interface	USB or serial
Weight	4.4 kg
Dimensions	24.5 x 35.6 x 28.6 – 37.4 cm
Delivery includes	The C-Quant basic unit Power supply unit Mains cable USB cable Dust cover Corrective lens set in wooden carrying case Instruction manual Installation CD

CE in accordance with Medical Device Directive 93/42/EEC



WWW.OCULUS.DE

OCULUS Optikgeräte GmbH  
Postfach • 35549 Wetzlar • GERMANY  
Tel. +49-641-2005-0 • Fax +49-641-2005-295  
E-Mail: export@oculus.de • www.oculus.de



OCULUS is certified by TÜV according to  
DIN EN ISO 13485/DIN EN ISO 9001

- OCULUS USA, sales@oculususa.com
- OCULUS Asia, info@oculus.hk
- OCULUS Czechia, oculus@oculus.cz
- OCULUS Iberia, info@oculus.es